

Offline Nano-ESI Phosphopeptide Analysis with Carbon, TiO₂ and ZrO₂ Wall-Coated Trap'nTips™


Christopher J. Toher¹, Adam W. Perala¹, Mukta M. Shukla², Ashok K. Shukla², Gary A. Valaskovic¹

¹New Objective, Inc., Woburn, MA ²Glygen, Inc., Columbia, MD

Introduction

Until recently, in-pipette sample preparation for nanobore LC-MS has been impeded by geometric constraints of sample preparation and transfer. Conventional practice has involved sample delivery from a pipette tip to an emitter, pressurized back-loading into a pipette, or coupling the pipette tip to a larger pipette tip¹.

The design of carbon, TiO₂ and ZrO₂ wall-coated pipette tips (Trap'nTips™, New Objective) facilitates phosphopeptide enrichment and manual aspiration through the pipette tip followed by expulsion of purified solution for immediate nanobore column injection or direct loading into an offline nanospray emitter. In addition to concentration, desalting, and MS signal enhancement, Trap'nTips coated with these analyte-specific substrates facilitate offline nanospray analysis by separating phosphopeptides from a tryptic β-casein digest.



Methods & Materials

Instrumentation and Components

- Ion-trap mass spectrometer (LCQ Deca™, Thermo Electron)
- Nanospray source (Digital PicoView® 150, New Objective, modified for offline analysis)
- A 0.5-10 µL Eppendorf® Single-Channel Research Pipette
- Sorbent-coated Trap'nTips™ containing carbon, TiO₂, and ZrO₂ sorbents (New Objective)
- GlassTips™ (BG12-94-4-CE) offline nanospray emitters (New Objective)

Sample Preparation


- Phosphopeptide Positive Control Set (Set P9615, Sigma) was prepared by diluting each vial with 1 mL HPLC-grade water / 0.1% formic acid.
- For each phosphopeptide, a 2 ng / µL solution was generated by diluting 10 µL of vial contents with 480 µL water / 0.1% formic acid solution.
- An overnight tryptic β-casein digest was performed, and the sample was diluted with water / 0.1% formic acid solution to generate a 1 pmol / µL solution.

Trap'nTip™ Conditioning and Sample Loading

- Trap'nTips were initially conditioned with 5 aspiration / expulsion cycles of HPLC-grade water.
- Analytes were loaded onto the Trap'nTip wall-coating through 10 aspiration / expulsion cycles of 10 µL ea
- Loaded samples were washed by 10 aspiration / expulsion cycles using 10 µL HPLC-grade water.
- Analytes were eluted from the Trap'nTip into a clean vial using 2 µL aqueous solution containing 50 mM NH₄HCO₃ and 50 mM TEA.

Offline NanoESI-MS Preparation

- 2 µL 50 mM TEA in methanol were added to the sample followed by immediate mixing and centrifugation.
- Each sample was loaded into the distal end of a GlassTip™ nanospray emitter (New Objective).
- All samples were analyzed via ESI-MS in negative-ion mode.




Results

GlassTips™ (New Objective) with a multi-layer platinum tip coating performed extremely well under high-pH conditions. A phosphopeptide standard was used to tune MS and to confirm peaks separated in the digest were, indeed, phosphopeptides. Carbon-, TiO₂-, and ZrO₂-coated Trap'nTips™ successfully separated phosphopeptides contained in the β -casein digest and phosphopeptide standard (Figure 3).

Conclusions

- For complex digest solutions, Carbon, TiO₂, and ZrO₂ wall-coated Trap'nTips™ successfully remove phosphopeptides for analysis via offline nanospray
- GlassTips™ with a multi-layer platinum-coated tip demonstrate robust performance in negative-ion mode
- Mass spectra of digest samples purified via carbon, TiO₂, and ZrO₂ Trap'nTips display exceptional signal-to-noise for phosphopeptide analysis



1. Toher, C.J.; Perala, A.W.; Shukla, A.K.; Valaskovic, G.A. "Sample Purification for Static Nanospray MS Using Wall-Coated Pipette Trap'nTips™". Poster presented at the American Society for Mass Spectrometry, San Antonio, 2005.